

Niagara Region
Public Health
1815 Sir Isaac Brock Way
P.O. Box 1052
Thorold, ON, L2V 0A2

905-688-8248 ext. 7590
Toll free: 1-888-505-6074 ext. 7590
niagararegion.ca/health

Sterilizer name:

Load number	Equipment identification	Cycle start time	Cycle end time	Cycle parameters			Malfunction or failure codes	Chemical indicator		Spore test strip *test strip run	Date (mm/dd/yyyy)	Initials
				Duration (minutes)	Temperature (°C/°F)	Pressure		External Pass/fail	Internal Pass/fail			

- Notes:
1. For failed internal/external chemical or biological indicator, ensure to follow internal policies and procedures; notify Niagara Region Public Health and initiate back-up plans.

2. Manufacturer's instructions for sterilizing equipment are to be onsite and easily accessible.

3. Always follow the manufacturer's instructions when reprocessing equipment.

4. Record book for monitoring each load placed in the sterilized must be maintained and onsite for a minimum of 1 year and an additional 2 years off-site

5. Document any preventative maintenance or repairs done on or to a sterilizer